



# CERTIFICATE OF ACCREDITATION

**ANSI-ASQ National Accreditation Board**  
500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

**Standards Calibration Laboratory**  
**Division of Global Gauge Corporation**  
**3200 Kettering Blvd.**  
**Moraine, OH 45439**


has been assessed by ANAB  
and meets the requirements of international standard

**ISO/IEC 17025:2005**

while demonstrating technical competence in the field of

**CALIBRATION**

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

AC-1122  
Certificate Number  
  
ANAB Approval

Certificate Valid: 05/12/2017-03/14/2019  
Version No. 007 Issued: 05/12/2017



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated January 2009).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Standards Calibration Laboratory
Division of Global Gauge Corporation

3200 Kettering Blvd.
Moraine, OH 45439
Tim McCormick / Wesley Bernard
937-254-3500

CALIBRATION

Valid to: March 14, 2019

Certificate Number: AC-1122

Length – Dimensional Metrology

Table with 4 columns: Parameter / Equipment, Range, Expanded Uncertainty of Measurement (+/-) 2, Reference Standard, Method and/or Equipment. Row 1: Thickness Standards, Up to 25.4 mm / Up to 1 in, (2 + 0.01L) μm / (80 + 6L) μin, SCL-001

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L represents the thickness measurement in appropriate unit of measurement stated above.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1122.

Signature of R.D. [unclear]
Vice President

